

# Instrument Maintenance Record

2018-19

DATE: / /  
PAGE NO.:

Sr.No.	Instrument Name	Batch Incharge	Problem	Action Taken	Issue Resolved	Outsource
1)	Hyogenizer	Mr. Maknikar K.P.	problem in grinding	plastic granules changed	Yes	Reminumbai Pvt Ltd Mumbai
2)	Spectrometer	Shinde M.H	Quartz Cuvettes Broken	replaced	Yes	ELICO Mumbai
3)	U.V. transilluminator	Maske P.S	Glassplate	Repaired	Self yes	—
4)	Cooling centrifuge	M.S. Sunvale	Machine	Repaired	Yes	Remi Sales Polghat
5)	Microscope	Mr. T. A. Patil	Servicing	Servicing	Self Yes	—

Madhusi  
Shinde M.H  
Incharge

NUE

H.O.D  
HEAD  
Department Of Biotechnology  
College of Computer Sci. & I.T  
Ambajogai Road, LATUR-413534

Teacher's Signature

Teacher's Signature

Instrument Maintenance Record  
2017-18

Sr No.	Instrument Name	Batch Incharge	Problem	Action Taken	Issue Resolved	Outsource
1)	Incubator	Mr. Malsnikar Kip	Temp Sensor not working	Repaired	Yes	Fourtech Pvt Ltd Pune
2)	Microscope	Mr. N.V. More	Microscope Binocular Probe	Repaired	Yes	ELICO Pune
3)	PH meter	Maske P.S	Probe	Repaired	Yes	ELICO Pune
4)	Laminar Air Flow	Mr. S.M. Incharje	servicing	servicing	Yes	—
5)	Incubator / oven	Mr. S.M. Incharje	servicing	servicing	Yes	—
6)	Heating mantle	Maske P.S	Plug	Repaired	Self	—

N.V.S.

H.O.D

HEAD

Department of Biotechnology  
College of Computer Sci. & I.T  
Ambajogai Road, LATUR-413551

*Kip*  
Incharge  
Mr. Malsnikar Kip

Instrument Maintenance Record


2015-16

DATE: / /  
PAGE NO.:

Sr No	Instrument Name	Batch Incharge	Problem	Action Taken	Issue Resolved	Outsource
1	pH meter	Mr. Malsnikar Kip	Electrode changed	changed	Yes	Atmaja Awarangabad
2	Heater bath	Mr. Malsnikar Kip	Not working Heating coil	Replaced	Yes	Atmaja Awarangabad
3	Incubator	MS. K.S. Gomare	Temp. fluctuation	Thermostat pocket repaired	Yes	Atmaja Awarangabad
4	PTC Lamp/air Flow	Ms B.T. Munde	servicing	servicing	Yes	Atmaja Awarangabad

MVS  
H.O.D

HEAD  
Department of Biotechnology  
College of Computer, Sci. & I.T.  
Ambajogal Road, Latur-413534

  
Incharge  
Mr. Malsnikar Kip

Teacher's Signature

Teacher's Signature

# Instrument Maintenance Record

DATE: / /  
PAGE NO.:

2014-15

SR.No	Instrument Name	Botch Incharge	problem	Action Taken	Issue Resolved	Signature
1	Double distillation assembly	Ms. B.T. Munde	coil	Replaced	Yes	Atmaja A'bad
2	Laminar Air Flow	Ms. B.T. Munde	UV Lamp	Replaced	Yes	Atmaja A'bad
3	Autoclave	Ms. S.M. Inkhuse	servicing	servicing	Yes	Atmaja A'bad
4	Vertical Laminar Air Flow	Ms. S.M. Inkhuse	Air Filter	servicing	Yes	Atmaja A'bad
5	Heating mantle	Ms. K.S. Gaware	servicing	coil changed	Yes	Atmaja A'bad

NVES

H.O.D

HEAD

Department Of Biotechnology  
College of Computer Sci. & I.T  
Ambajogai Road, LATUR-413534

*(Signature)*  
Incharge

Teacher's Signature

Teacher's Signature

Instrument Maintenance

Record

2017-18

DATE: / /  
PAGE NO.:

Sr No.	Instrument Name	Batch Incharge
1)	Incubator	Mr. Malnikas Kip
2)	Microscope	Mr. N.V. More
3)	pH meter	Maske P.S
4)	Laminar Air flow	Mr. S.M. Incharge
5)	Incubator / oven	Mr. S.M. Incharge
6)	Heating mantle	Maske P.S

*Malnikas Kip*  
Incharge  
Mr. Malnikas Kip.

Teacher's Signature

DATE: / /  
PAGE NO.:

Problem	Action Taken	Issue Resolved	Outsource
Temp Sensor not working	Repaired	Yes	Fourtech Pvt Ltd pune
Monoculor Biorad probe	Repaired	Yes	ELICO Pune
servicing	servicing	Yes	Balesh poudre
servicing	servicing	Yes	(-11-)
Plug	Repaired	Self	-
		yes	-

*Malnikas Kip*  
H.O.D

HEAD  
Department of Biotechnology  
College of Computer Sci. & I.T  
Ambajogai Road, Latur-41353

Teacher's Signature

Scanned by CamScanner

Instrument Maintenance Record  
2016-17

DATE: / /  
PAGE NO.:

SR No	Instrument Name	Batch Incharge
1	Cooling centrifuge	Mr. Makrikar, K.P.
2	Staking incubator	Ms. Karmal S. Gore
3	Water bath Staker	Ms. Karmal S. Gore
4	Heating Mantle	Mr. V.M. Navisetti
5	Was burner & pipes	Mr. V. Pobil.

*[Signature]*  
Incharge  
Mr. Karmal Gore

Teacher's Signature

DATE: / /  
PAGE NO.:

Problem	Action Taken	Issue Resolved	Outsource
cooling problem	repaired	Yes	Remi Det. Lab pane
temp. maintenance	Repaired	Yes	Remi put lab pane
coil problem	Repaired	Yes	Remi put lab pane
Repaired	Repaired	Yes	Remi put lab pane
Repaired	Severing	Yes	Fluoresce lab pane

*[Signature]*  
H.O.D  
HEAD  
Department of Biotechnology  
College of Computer Sci. & I.T  
Ambajogai Road, Latur-430507

Teacher's Signature

Scanned by CamScanner

Instrument Maintenance Record

2015-16

DATE: / /  
PAGE NO.:

SRNO	Instrument Name	Batch Incharge
1	pH meter	Mr. Maknikar K.P.
2	Heater bath	Mr. Maknikar K.P.
3	Incubator	MS. K.S. Gomare
4	PTC Laminar Air Flow	Mr. B.T. Munde

~~Mr. Maknikar K.P.~~  
Incharge  
Mr. Maknikar K.P.

Teacher's Signature

DATE: / /  
PAGE NO.:

Problem	Action Taken	Issue Resolved	Outsource
Electrode changed	changed	Yes	Atmaja Arranged
Not working Heating coil	Replaced	Yes	Atmaja Arranged
Temp. fluctuation	Thermostat pocket repaired	Yes	Atmaja Arranged
servicing	servicing	Yes	Atmaja Arranged

NUS  
H.O.D

HEAD  
Department Of Biotechnology  
College of Computer Sci. & I.T.  
Ambajogal Road, LATUR-41353

Teacher's Signature

# Instrument Maintenance Record

2014-15

DATE: / /  
PAGE NO.:



DATE: / /  
PAGE NO.:

SR.NO	Instrument Name	Batch Incharge
1	Double distillation assembly	Ms. B.T. Munde
2	Laminar Air flow	Ms. B.T. Munde
3	Autoclave	Ms. S.M. Incharge
4	Vertical laminar Air flow	Ms. S.M. Incharge
5	Heating mantle	Ms. K.S. Gomare

Problem	Action Taken	Issue Resolved	Observed
coil	Replaced	Yes	Atmaja Abad
UV lamp	Replaced	Yes	Atmaja Abad
servicing	servicing	Yes	Atmaja Abad
Air filter	servicing	Yes	Atmaja Abad
servicing	coil changed	Yes	Atmaja Abad

Ms. S.M.  
Incharge

Atmaja  
H.O.D  
HEAD  
Department Of Biotechnology  
College of Computer Sci. & I.T.  
Ambajogai Road, LATUR-413531